



Issue Date: 13-05-2025

This is to certify that verification of "Essae Milk Analyzer MA-915ABS" equipment (Model: MA-915ABS, Sr. no.: B2400115049) submitted by M/s Essae-Teraoka Pvt. Ltd., Koramangala, Bengaluru was carried out at NDDB CALF Ltd., Anand, Gujarat.

The equipment exhibited satisfactory performance for the claimed accuracy of Fat ($\pm 0.1\%$) and SNF ($\pm 0.2\%$) in the working range (Fat: 0.5 to 15%, SNF: 3 to 15%) (as per report no. NCL/Project/2025-26/02).

The validity of the report & certificate is for three years from the date of issue and it is subjected to the consistent manufacturing of the equipment as submitted for verification.

Authorized Signatory,

(Director)

The equipment has been verified with proper calibration under standard laboratory conditions. The legal responsibility for compliance of the equipment is solely that of the holder of this Certificate, and NDDB CALF LIMITED does not warrant that the holder of certificate at all times complies with the provisions of the said certificate and report.